

## ***Experimental Envelope for the SBC 19BM Facility*** **(September 2001)**

The commissioning envelope described below does not define the limits of the beamline capabilities but describes conditions for crystallographic experiments conducted successfully at the beamline.

### **Sample Characteristics**

- Crystals of macromolecules known to diffract (10 - 500 kDa/AU)
- Unit cell dimension: longest  $\leq 600 \text{ \AA}$
- Number of heavy atom sites validated
- Anomalous scatterers within the beamline energy limits (results demonstrated for Se, Co, Fe, Pt, Hg, S, Xe, I; currently maximum number of sites – 12/A.U.)
- Crystal size: 0.05 – 1.0 mm
- Cryofreezing protocol established
- Mosaicity  $\leq 2^\circ$

### **Sample Preparation Capabilities**

- Crystallization facility at the SBC available (special arrangements on request)
- On-site sample freezing in liquid nitrogen, nitrogen cold stream, or liquid propane
- Cold room available ( $4^\circ \text{ C}$ )
- Wet lab and modest lab equipment available (users bring their own chemicals or arrange for shipment to ANL)

### **X-ray Beam Characteristics 19ID**

- Energy: 6.5 - 13 keV (special arrangements request)
- Energy band width:  $1.8 \times 10^{-4}$
- Collimator size: 0.2 mm x 0.15mm
- Flux: As needed, limited by sample degradation (max.  $6 \times 10^{12}$  x-ray Ph/sec/mm<sup>2</sup>)
- Beamstop – 1 mm, designed for MAD phasing to limit fluorescence interference

### **Data Acquisition Conditions**

- Exposure: 1 - 180 sec per frame
- Sample temperature:  $\geq 94 \text{ K}$  (Oxford)
- Detector-sample distance:  $\geq 120 - \leq 960 \text{ mm}$
- Detector 2 theta angle:  $-5^\circ - +38^\circ$
- Detector offset  $-100 - +400 \text{ mm}$
- Kappa goniostat operational (motor-driven phi rotation and z-translation, manual x- and y-translation)
- Omega angle:  $\pm 120^\circ$  continuous scan range
- Omega angle drive rate:  $\leq 10^\circ/\text{sec}$
- Oscillation angle: fine or wide slicing ( $0.1 - 4.0^\circ$ )
- Mounting pin type: Yale, Harvard, Hampton, (special arrangements on request)

- MAD experiments: presently with operator assistance

### **Special Instrumentation**

- **Fluorescence detectors**
  - Amptek Model CR100, energy resolution  $\sim 350$  eV at 12 keV, 1-2K cps for linear detection
  - BICRON, poor energy resolution, but 30-40K cps for linear detection
- **X-ray area detector (SBC-1 and SBC-2) characteristics**
  - Nine-element CCD
  - Surface area: 210 x 210 mm
  - Pixel size: 0.068 mm
  - Unbinned images: 3072 x 3072 pixels (18.5 Mb image size, 2.9 sec deadtime)
  - Binned images: 1536 x 1536 pixels (4.5 Mb image size, 1.6 sec deadtime)
  - Data acquisition rate – up to approximately 30 images/min

### **Data Processing and Management**

- HKL2000 v0.96.511, d\*TREK v.7.1/v.7.2
- Automated data management (transfer and archiving)
- Data transfer rates 3.5-4.5 Mb/sec
- Data archive rates approximately 2.8 Mb/sec
- SGI Challenge computer (2 x 200 MHz processors) with 50 Gb disk space
- One PC LINUX workstation with dual processors (450 MHz), two with dual processors (1 GHz), two with dual processors (700 MHz) and two 220 Gb RAID5 disk storage spaces
- Archive data to user-provided DLT (20/35/70 Gb capacity; compressed) or DAT (12/8/4 Gb capacity) tape

### **Personnel and Administration**

- Independent investigator agreement between user's institution and the APS
- Completion of proposal review form and safety review form
- Identification of spokesperson designated by the principal investigator
- Completion of DOE-approved radiation safety training (available at the APS)
- Completion of ANL-approved beamline training (available at the SBC)
- Current SBC support team assigned to each user group

User support: day one: full support 9 am to 9 pm, subsequent days: full support during regular hours; weekends and off-hour support only by special arrangement